

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3917

Name: RITTER 1  
Sterilizer ID: V2178198  
10/18/2023 09:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.8 F 0.0  
2:00 144.1 F 1.8  
4:00 150.8 F 4.6  
6:00 176.7 F 8.7  
8:00 202.8 F 13.5  
10:00 220.9 F 6.6  
12:00 232.0 F 11.4  
14:00 233.0 F 7.7  
16:00 245.4 F 12.8  
18:00 243.6 F 12.0  
20:00 256.5 F 18.6  
22:00 267.9 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.6  
1:00 271.8 F 28.5  
2:00 272.6 F 28.8  
3:00 272.4 F 29.1  
4:00 271.5 F 28.7

Min 271.4 F 28.5  
Max 273.0 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.5 F 0.0  
5:00 237.2 F 0.0  
10:00 217.3 F 0.0  
15:00 193.0 F 0.0  
20:00 191.8 F 0.0  
25:00 197.8 F 0.0  
30:00 188.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 23:57  
STERILIZING: 4:00  
VENTING: 3:24  
DRYING: 30:00  
TOTAL CYCLE: 01:02:19  
CYCLE COMPLETE

10/18/2023 10:16 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_