
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000808
11:56 23/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:22
PRESSURIZING 6:12
STERILIZING 7:53
136.4 C 222kPa 7:53
136.4 C 223kPa 8:23
136.0 C 219kPa 8:53
136.3 C 222kPa 9:23
136.4 C 220kPa 9:53
136.5 C 223kPa 10:23
136.4 C 223kPa 10:53
136.3 C 222kPa 11:23
136.3 C 222kPa 11:23
VENTING 11:23
AIR DRYING 12:35
CYCLE COMPLETE 42:35

Digital Signature #
3ADC619D6B9607EF
