
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000088
11:09 13/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:24
PRESSURIZING 3:25
STERILIZING 4:26
136.0 C 220kPa 4:26
135.9 C 222kPa 4:56
136.4 C 224kPa 5:26
136.4 C 223kPa 5:56
136.0 C 223kPa 6:26
136.2 C 222kPa 6:56
136.4 C 225kPa 7:26
136.1 C 220kPa 7:56
136.1 C 220kPa 7:56
VENTING 7:56
AIR DRYING 8:48
CYCLE COMPLETE 8:59

Digital Signature #
8AB76D5887274363
