

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2491

Name: RITTER2  
Sterilizer ID: V2313228  
06/09/2023 07:22 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.1 F 0.0  
2:00 78.2 F 0.0  
4:00 100.5 F 0.0  
6:00 114.8 F 0.0  
8:00 148.1 F 1.0  
10:00 173.6 F 4.0  
12:00 195.2 F 7.5  
14:00 203.9 F 11.5  
16:00 219.2 F 5.0  
18:00 227.3 F 9.4  
20:00 238.1 F 11.7  
22:00 239.2 F 11.2  
24:00 236.6 F 9.4  
26:00 249.9 F 15.7  
28:00 260.8 F 22.0  
30:00 271.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.3 F 29.8  
2:00 272.1 F 29.1  
3:00 271.9 F 29.6  
4:00 272.4 F 29.5

Min 271.4 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.9 F 0.0  
5:00 201.4 F 0.0  
10:00 208.9 F 0.0  
15:00 218.0 F 0.0  
20:00 220.1 F 0.0  
25:00 220.0 F 0.0  
30:00 231.4 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 30:34  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:08:34  
CYCLE COMPLETE

06/09/2023 08:30 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_