

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1382

Name: RITTER 1  
Sterilizer ID: V2178198  
11/18/2021 09:30 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.2 F 0.0  
2:00 148.6 F 1.0  
4:00 158.7 F 3.9  
6:00 181.4 F 7.8  
8:00 202.3 F 12.7  
10:00 222.4 F 5.5  
12:00 231.8 F 10.4  
14:00 233.5 F 7.7  
16:00 244.0 F 12.7  
18:00 241.8 F 11.2  
20:00 254.5 F 17.7  
22:00 265.9 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.5 F 29.1  
2:00 272.0 F 28.5  
3:00 271.3 F 28.4  
4:00 272.6 F 29.1

Min 271.2 F 28.2  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 197.6 F 0.0  
10:00 203.2 F 0.0  
15:00 204.5 F 0.0  
20:00 214.5 F 0.0  
25:00 216.9 F 0.0  
30:00 214.4 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 23:44  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:02:03  
CYCLE COMPLETE

11/18/2021 10:39 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02