
Midmark M11 - v1.0.5
Total Cycles: 1668

Name: RITTER2
Sterilizer ID: V2313228
11/02/2022 10:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.5 F 0.0
2:00 174.9 F 2.8
4:00 198.0 F 6.6
6:00 217.9 F 11.6
8:00 219.0 F 6.1
10:00 227.1 F 11.6
12:00 231.0 F 8.4
14:00 240.6 F 11.2
16:00 244.4 F 13.1
18:00 256.8 F 19.7
20:00 268.0 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.3
1:00 272.3 F 29.6
2:00 271.8 F 29.7
3:00 272.1 F 29.4
4:00 271.9 F 29.9

Min 271.4 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 217.3 F 0.0
10:00 238.6 F 0.0
15:00 232.9 F 0.0
20:00 227.2 F 0.0
25:00 226.2 F 0.0
30:00 225.8 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:41
CYCLE COMPLETE

11/02/2022 11:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____