
Midmark M11 - v1.0.5
Total Cycles: 1781

Name: RITTER2
Sterilizer ID: V2313228
11/28/2022 04:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 211.1 F 0.0
2:00 202.3 F 4.6
4:00 215.2 F 10.9
6:00 225.4 F 7.7
8:00 237.1 F 11.5
10:00 243.1 F 13.1
12:00 244.3 F 12.9
14:00 257.7 F 20.2
16:00 269.9 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.1
1:00 272.3 F 29.5
2:00 272.3 F 29.8
3:00 272.3 F 29.3
4:00 272.2 F 29.8

Min 271.6 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 232.6 F 0.0
10:00 233.8 F 0.0
15:00 231.0 F 0.0
20:00 230.2 F 0.0
25:00 221.7 F 0.0
30:00 228.1 F 0.0
Drying Complete

FILLING: 3:38
HEATING: 16:52
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:41
CYCLE COMPLETE

11/28/2022 05:35 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____