

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001771  
15:13 02/05/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:22  
STERILIZING 3:26  
136.0 C 220kPa 3:26  
136.1 C 223kPa 3:56  
136.0 C 222kPa 4:26  
136.2 C 226kPa 4:56  
136.2 C 224kPa 5:26  
136.4 C 226kPa 5:56  
136.3 C 225kPa 6:26  
136.1 C 223kPa 6:56  
136.1 C 223kPa 6:56  
VENTING 6:56  
AIR DRYING 7:46  
CYCLE COMPLETE 8:11  
-----

Digital Signature #  
409E9B3E3D2513E1  
-----