
Midmark M11 - v1.0.5
Total Cycles: 3762

Name: RITTER2
Sterilizer ID: V2313228
07/10/2024 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 70.6 F 0.0
4:00 76.9 F 0.0
6:00 106.6 F 0.1
8:00 146.0 F 0.7
10:00 169.7 F 3.4
12:00 189.4 F 6.4
14:00 203.2 F 9.7
16:00 216.6 F 13.1
18:00 222.4 F 6.5
20:00 233.5 F 10.5
22:00 232.7 F 8.8
24:00 241.3 F 12.1
26:00 239.2 F 10.4
28:00 250.9 F 16.2
30:00 261.2 F 22.0
32:00 270.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.4 F 29.8
2:00 271.9 F 29.5
3:00 271.9 F 29.2
4:00 272.7 F 29.4

Min 270.9 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.8 F 0.0
5:00 186.2 F 0.0
10:00 178.9 F 0.0
15:00 175.7 F 0.0
20:00 173.9 F 0.0
25:00 172.3 F 0.0
30:00 172.7 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:47
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

07/10/2024 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____