
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001606
13:44 11/22/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:39
PRESSURIZING 6:26
STERILIZING 8:00
136.4 C 222kPa 8:00
136.3 C 221kPa 8:30
136.3 C 222kPa 9:00
136.4 C 222kPa 9:30
136.4 C 220kPa 10:00
136.2 C 219kPa 10:30
136.4 C 220kPa 11:00
136.5 C 219kPa 11:30
VENTING 11:30
AIR DRYING 12:33
CYCLE COMPLETE 24:00

Digital Signature #
768C29A6FE28AC01
