
Midmark M11 - v1.0.5
Total Cycles: 1021

Name: RITTER2
Sterilizer ID: V2313228
04/22/2022 10:38 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 142.5 F 0.0
2:00 143.1 F 0.0
4:00 159.3 F 0.2
6:00 187.7 F 2.9
8:00 211.7 F 8.6
10:00 226.2 F 7.9
12:00 237.2 F 11.8
14:00 238.6 F 10.1
16:00 237.2 F 9.1
18:00 252.6 F 16.6
20:00 264.9 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.5
1:00 272.3 F 28.7
2:00 273.0 F 29.3
3:00 271.8 F 28.6

Min 271.6 F 28.3
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
Dry Cycle Aborted

FILLING: 1:18
HEATING: 21:43
STERILIZING: 3:00
VENTING: 2:49
DRYING: 1:32
TOTAL CYCLE: 00:30:23

04/22/2022 11:08 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____