

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1849

Name: RITTER2  
Sterilizer ID: V2313228  
12/15/2022 01:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.5 F 0.0  
2:00 171.5 F 0.7  
4:00 189.1 F 5.0  
6:00 201.6 F 10.7  
8:00 218.2 F 5.3  
10:00 229.5 F 11.0  
12:00 231.9 F 8.3  
14:00 239.6 F 10.7  
16:00 246.2 F 14.0  
18:00 259.0 F 20.9  
20:00 270.5 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.2  
1:00 272.4 F 29.7  
2:00 271.4 F 29.2  
3:00 272.3 F 29.7  
4:00 271.5 F 29.1  
  
Min 271.4 F 29.0  
Max 272.5 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.7 F 0.0  
5:00 206.5 F 0.0  
10:00 218.3 F 0.0  
15:00 234.4 F 0.0  
20:00 234.4 F 0.0  
25:00 224.2 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 20:47  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:05  
CYCLE COMPLETE

12/15/2022 02:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_