
Midmark M11 - v1.0.5
Total Cycles: 1639

Name: RITTER2
Sterilizer ID: V2313228
10/21/2022 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.7 F 0.0
2:00 175.3 F 3.2
4:00 188.1 F 7.6
6:00 208.3 F 13.1
8:00 222.1 F 8.1
10:00 234.8 F 12.3
12:00 238.5 F 11.3
14:00 238.1 F 10.1
16:00 252.2 F 17.1
18:00 264.1 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.3 F 29.6
2:00 271.3 F 29.2
3:00 272.4 F 29.8
4:00 271.6 F 29.1

Min 271.3 F 29.0
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 214.3 F 0.0
10:00 228.4 F 0.0
15:00 238.9 F 0.0
20:00 230.3 F 0.0
25:00 225.7 F 0.0
30:00 213.1 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 19:57
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

10/21/2022 03:21 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____