
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001550
12:38 30/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:47
STERILIZING 3:40
136.0 C 220kPa 3:40
135.9 C 222kPa 4:10
136.4 C 223kPa 4:40
136.2 C 223kPa 5:10
136.4 C 223kPa 5:40
136.3 C 222kPa 6:10
136.2 C 223kPa 6:40
136.4 C 225kPa 7:10
136.4 C 225kPa 7:10
VENTING 7:11
AIR DRYING 8:02
CYCLE COMPLETE 8:20

Digital Signature #
181987A56EE3E154
