
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001682
10:43 11/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:53
STERILIZING 4:07
136.0 C 220kPa 4:07
136.1 C 222kPa 4:37
135.7 C 219kPa 5:07
136.3 C 224kPa 5:37
136.4 C 225kPa 6:07
136.4 C 224kPa 6:37
136.0 C 222kPa 7:07
136.3 C 222kPa 7:37
136.3 C 222kPa 7:37
VENTING 7:37
AIR DRYING 8:28
CYCLE COMPLETE 9:15

Digital Signature #
F8A55A673C081676
