

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002363  
10:42 06/12/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:25  
STERILIZING 3:24  
136.3 C 224kPa 3:24  
136.4 C 223kPa 3:54  
136.4 C 223kPa 4:24  
136.4 C 225kPa 4:54  
136.5 C 226kPa 5:24  
136.4 C 223kPa 5:54  
136.4 C 223kPa 6:24  
136.4 C 226kPa 6:54  
136.4 C 226kPa 6:54  
VENTING 6:54  
AIR DRYING 7:44  
CYCLE COMPLETE 37:44

-----  
Digital Signature #  
7EC453B13946E28E  
-----