
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005747
9:49 13/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:35
STERILIZING 3:38
136.5 C 222kPa 3:38
136.3 C 225kPa 4:08
136.1 C 223kPa 4:38
136.4 C 221kPa 5:08
136.4 C 224kPa 5:38
136.3 C 221kPa 6:08
136.2 C 221kPa 6:38
136.4 C 225kPa 7:08
136.4 C 225kPa 7:08
VENTING 7:08
AIR DRYING 8:01
CYCLE COMPLETE 8:05

Digital Signature #
C447597ABA7CABEF
