
Midmark M11 - v1.0.0
Total Cycles: 1540

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2022 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 204.7 F 0.0
2:00 197.6 F 4.2
4:00 211.7 F 9.5
6:00 224.2 F 6.5
8:00 234.3 F 10.9
10:00 233.8 F 8.1
12:00 240.7 F 10.9
14:00 245.9 F 13.4
16:00 258.4 F 20.2
18:00 269.8 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.1 F 28.9
2:00 272.4 F 29.4
3:00 271.5 F 28.6
4:00 272.5 F 29.3

Min 271.3 F 28.5
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.0 F 0.0
5:00 226.9 F 0.0
10:00 237.3 F 0.0
15:00 231.4 F 0.0
20:00 225.7 F 0.0
25:00 225.8 F 0.0
30:00 229.0 F 0.0

Drying Complete

FILLING: 2:06
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

01/19/2022 03:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____