
Midmark M11 - v1.0.0
Total Cycles: 4174

Name: RITTER 1
Sterilizer ID: V2178198
01/03/2024 11:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.3 F 0.0
2:00 184.4 F 2.5
4:00 184.6 F 6.1
6:00 199.3 F 10.5
8:00 218.4 F 5.4
10:00 224.8 F 8.1
12:00 234.3 F 12.5
14:00 235.3 F 9.3
16:00 240.3 F 10.1
18:00 246.3 F 13.5
20:00 257.7 F 19.6
22:00 268.1 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.3 F 28.9
2:00 271.3 F 28.6
3:00 272.2 F 28.8
4:00 271.2 F 28.4

Min 271.2 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 228.3 F 0.0
10:00 230.4 F 0.0
15:00 225.6 F 0.0
20:00 222.8 F 0.0
25:00 222.4 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

01/03/2024 12:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____