

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3867

Name: RITTER 1  
Sterilizer ID: V2178198  
10/04/2023 07:36 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 171.8 F 0.0  
2:00 180.9 F 3.9  
4:00 204.1 F 9.3  
6:00 224.2 F 6.2  
8:00 233.7 F 10.5  
10:00 232.5 F 7.6  
12:00 242.1 F 10.9  
14:00 247.0 F 13.6  
16:00 260.4 F 20.8  
18:00 272.1 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.6  
1:00 272.0 F 28.8  
2:00 272.3 F 29.1  
3:00 272.5 F 29.1  
4:00 272.3 F 29.1  
  
Min 271.8 F 28.5  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.4 F 0.0  
5:00 234.0 F 0.0  
10:00 224.3 F 0.0  
15:00 223.2 F 0.0  
20:00 181.9 F 0.0  
25:00 178.7 F 0.0  
30:00 177.8 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 19:04  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

10/04/2023 08:34 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_