
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006205
8:34 30/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:21
STERILIZING 4:27
136.0 C 220kPa 4:27
136.4 C 223kPa 4:57
136.4 C 223kPa 5:27
136.4 C 221kPa 5:57
136.4 C 223kPa 6:27
136.4 C 221kPa 6:57
136.3 C 219kPa 7:27
136.4 C 219kPa 7:57
136.4 C 219kPa 7:57
VENTING 7:57
AIR DRYING 8:45
CYCLE COMPLETE 34:39

Digital Signature #
0DB31344556C57CA
