
Midmark M11 - v1.0.0
Total Cycles: 5376

Name: RITTER 1
Sterilizer ID: V2178198
02/28/2025 09:33 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 89.2 F 0.4
4:00 94.3 F 2.0
6:00 125.2 F 4.4
8:00 153.6 F 7.3
10:00 175.8 F 10.7
12:00 194.7 F 3.7
14:00 213.6 F 4.6
16:00 221.5 F 8.0
18:00 232.2 F 11.7
20:00 229.2 F 6.7
22:00 241.2 F 11.7
24:00 237.2 F 9.1
26:00 250.2 F 15.3
28:00 261.3 F 21.6
30:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.4 F 28.8
2:00 271.8 F 28.6
3:00 272.1 F 28.6
4:00 272.1 F 28.8

Min 271.5 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.1 F 0.0
5:00 190.3 F 0.0
10:00 189.1 F 0.0
15:00 197.4 F 0.0
20:00 189.8 F 0.0
25:00 191.7 F 0.0
30:00 192.7 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 30:37
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:08:58
CYCLE COMPLETE

02/28/2025 10:41 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____