

-----  
Midmark M11 - v1.0.5  
Total Cycles: 369

Name: RITTER2  
Sterilizer ID: V2313228  
09/16/2021 07:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 187.5 F 0.0  
2:00 179.1 F 2.4  
4:00 191.3 F 6.7  
6:00 210.1 F 12.0  
8:00 220.8 F 7.5  
10:00 234.6 F 13.1  
12:00 238.0 F 11.0  
14:00 237.9 F 9.7  
16:00 251.6 F 16.5  
18:00 263.0 F 23.3  
20:00 271.6 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.1  
1:00 271.8 F 29.5  
2:00 272.1 F 29.5  
3:00 272.3 F 29.3  
4:00 271.9 F 29.2

Min 271.4 F 29.0  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.4 F 0.0  
5:00 227.4 F 0.0  
10:00 227.1 F 0.0  
15:00 215.3 F 0.0  
20:00 208.8 F 0.0  
25:00 228.4 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 20:00  
STERILIZING: 4:00  
VENTING: 3:01  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

09/16/2021 08:50 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02