
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002912
11:18 30/12/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:42
STERILIZING 3:44
136.0 C 220kPa 3:44
136.4 C 224kPa 4:14
136.2 C 220kPa 4:44
136.2 C 221kPa 5:14
136.5 C 224kPa 5:44
136.4 C 223kPa 6:14
136.1 C 220kPa 6:44
136.3 C 222kPa 7:14
136.3 C 222kPa 7:14
VENTING 7:14
AIR DRYING 8:08
CYCLE COMPLETE 8:20

Digital Signature #
AF77ACE451A4826D
