

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3928

Name: RITTER 1  
Sterilizer ID: V2178198  
10/20/2023 09:39 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.1 F 0.0  
2:00 134.8 F 0.0  
4:00 137.7 F 2.1  
6:00 165.4 F 5.0  
8:00 189.6 F 9.1  
10:00 206.8 F 13.3  
12:00 215.7 F 6.2  
14:00 226.4 F 10.6  
16:00 230.2 F 7.6  
18:00 240.7 F 11.9  
20:00 238.2 F 9.9  
22:00 250.4 F 15.8  
24:00 261.2 F 21.9  
26:00 271.1 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.4 F 29.3  
2:00 271.5 F 28.7  
3:00 271.7 F 29.0  
4:00 272.4 F 29.3

Min 271.0 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.2 F 0.0  
5:00 229.1 F 0.0  
10:00 234.1 F 0.0  
15:00 229.8 F 0.0  
20:00 230.0 F 0.0  
25:00 223.1 F 0.0  
30:00 217.4 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 26:44  
STERILIZING: 4:00  
VENTING: 3:42  
DRYING: 30:00  
TOTAL CYCLE: 01:05:19  
CYCLE COMPLETE

10/20/2023 10:44 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_