
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001647
15:59 05/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:10
STERILIZING 3:12
136.0 C 222kPa 3:12
136.3 C 225kPa 3:42
136.4 C 226kPa 4:12
136.2 C 222kPa 4:42
136.4 C 225kPa 5:12
136.4 C 225kPa 5:42
136.3 C 225kPa 6:12
136.3 C 224kPa 6:42
136.3 C 224kPa 6:42
VENTING 6:42
AIR DRYING 7:40
CYCLE COMPLETE 7:50

Digital Signature #
4A3C4F932C5BD3AF
