

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007621  
10:00 02/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:47  
PRESSURIZING 3:48  
STERILIZING 4:52  
136.5 C 224kPa 4:52  
136.3 C 223kPa 5:22  
136.1 C 220kPa 5:52  
136.4 C 224kPa 6:22  
136.3 C 224kPa 6:52  
136.3 C 224kPa 7:22  
136.4 C 224kPa 7:52  
136.3 C 220kPa 8:22  
136.3 C 220kPa 8:22  
VENTING 8:22  
AIR DRYING 9:17  
CYCLE COMPLETE 9:47  
-----

Digital Signature #  
CF42A7A2FD459BC3  
-----