
Midmark M11 - v1.0.5
Total Cycles: 226

Name: RITTER2
Sterilizer ID: V2313228
07/20/2021 08:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.9 F 0.0
2:00 103.6 F 0.6
4:00 113.1 F 2.3
6:00 137.1 F 5.3
8:00 165.0 F 8.9
10:00 190.4 F 13.0
12:00 208.8 F 5.5
14:00 223.5 F 9.8
16:00 229.4 F 9.5
18:00 237.4 F 11.8
20:00 237.8 F 10.2
22:00 250.3 F 16.3
24:00 261.1 F 22.6
26:00 271.4 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.6
1:00 272.7 F 30.2
2:00 271.7 F 29.5
3:00 272.8 F 30.3
4:00 272.1 F 29.6

Min 271.7 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.2 F 0.0
5:00 199.6 F 0.0
10:00 206.1 F 0.0
15:00 210.4 F 0.0
20:00 212.2 F 0.0
25:00 215.4 F 0.0
30:00 223.0 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 26:39
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:04:47
CYCLE COMPLETE

07/20/2021 09:39 AM
ACCEPTED: X
REJECTED:
Operator: U01