
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 002750
8:46 30/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:22
STERILIZING 4:28
136.0 C 220kPa 4:28
136.4 C 223kPa 4:58
136.4 C 223kPa 5:28
136.3 C 223kPa 5:58
136.1 C 221kPa 6:28
136.2 C 220kPa 6:58
136.2 C 221kPa 7:28
136.2 C 221kPa 7:58
136.2 C 221kPa 7:58
VENTING 7:58
AIR DRYING 8:50
CYCLE COMPLETE 9:17

Digital Signature #
59702F697095D52B
