

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2071

Name: RITTER 1  
Sterilizer ID: V2178198  
06/29/2022 07:28 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.8 F 0.0  
2:00 73.4 F 0.0  
4:00 78.6 F 0.0  
6:00 107.5 F 0.8  
8:00 142.4 F 2.3  
10:00 169.8 F 4.2  
12:00 212.3 F 7.4  
14:00 220.0 F 11.1  
16:00 224.0 F 4.5  
18:00 230.7 F 7.7  
20:00 240.7 F 11.9  
22:00 234.6 F 8.2  
24:00 245.1 F 12.2  
26:00 244.0 F 12.0  
28:00 255.7 F 18.1  
30:00 266.1 F 24.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.1  
1:00 272.4 F 28.9  
2:00 271.9 F 28.2  
3:00 271.5 F 28.4  
4:00 272.4 F 28.8  
  
Min 271.2 F 28.0  
Max 272.5 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
5:00 201.1 F 0.0  
10:00 188.3 F 0.0  
15:00 217.9 F 0.0  
20:00 238.1 F 0.0  
25:00 229.9 F 0.0  
30:00 224.5 F 0.0  
Drying Complete

FILLING: 0:44  
HEATING: 31:48  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 01:09:38  
CYCLE COMPLETE

06/29/2022 08:37 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_