
Midmark M11 - v1.0.0
Total Cycles: 3068

Name: RITTER 1
Sterilizer ID: V2178198
04/03/2023 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.8 F 0.0
2:00 195.3 F 3.3
4:00 210.5 F 8.3
6:00 222.6 F 10.3
8:00 231.5 F 9.4
10:00 238.2 F 9.1
12:00 243.9 F 13.1
14:00 243.3 F 12.1
16:00 256.6 F 19.1
18:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.3 F 28.9
2:00 272.4 F 29.4
3:00 271.9 F 28.6
4:00 272.7 F 29.4

Min 271.6 F 28.5
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 228.9 F 0.0
10:00 238.6 F 0.0
15:00 232.1 F 0.0
20:00 225.8 F 0.0
25:00 225.6 F 0.0
30:00 225.2 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:10
CYCLE COMPLETE

04/03/2023 02:07 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____