

-----  
Midmark M11 - v1.0.0  
Total Cycles: 847

Name: RITTER 1  
Sterilizer ID: V2178198  
05/27/2021 03:21 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.3 F 0.0  
2:00 169.0 F 2.2  
4:00 179.1 F 5.2  
6:00 196.7 F 9.0  
8:00 212.1 F 12.9  
10:00 220.5 F 6.5  
12:00 234.1 F 10.9  
14:00 237.4 F 7.7  
16:00 242.7 F 12.8  
18:00 241.8 F 11.1  
20:00 253.0 F 16.9  
22:00 263.2 F 22.9  
24:00 272.6 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.4  
1:00 272.7 F 29.2  
2:00 272.4 F 28.8  
3:00 271.8 F 28.3  
4:00 272.2 F 28.8

Min 271.7 F 28.3  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.1 F 0.0  
5:00 203.4 F 0.0  
10:00 221.2 F 0.0  
15:00 216.3 F 0.0  
20:00 205.5 F 0.0  
25:00 212.2 F 0.0  
30:00 206.3 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 24:28  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

05/27/2021 04:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_