
Midmark M11 - v1.0.0
Total Cycles: 1616

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2022 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 191.6 F 3.8
4:00 198.8 F 8.4
6:00 218.2 F 10.6
8:00 227.3 F 8.9
10:00 236.8 F 10.3
12:00 240.8 F 11.9
14:00 239.7 F 10.6
16:00 253.2 F 17.4
18:00 265.1 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 29.3
2:00 271.5 F 28.5
3:00 272.7 F 29.5
4:00 272.2 F 29.0

Min 271.5 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.7 F 0.0
5:00 216.7 F 0.0
10:00 230.7 F 0.0
15:00 235.5 F 0.0
20:00 227.6 F 0.0
25:00 220.3 F 0.0
30:00 218.1 F 0.0
Drying Complete

FILLING: 2:06
HEATING: 19:46
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

02/10/2022 12:50 PM
ACCEPTED: X
REJECTED:
Operator: U02