
Midmark M11 - v1.0.0
Total Cycles: 2258

Name: RITTER 1
Sterilizer ID: V2178198
08/29/2022 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 176.2 F 0.3
4:00 194.0 F 1.1
6:00 213.6 F 4.4
8:00 229.5 F 9.1
10:00 239.9 F 10.7
12:00 239.6 F 11.1
14:00 237.2 F 9.1
16:00 239.2 F 9.3
18:00 250.0 F 15.1
20:00 262.5 F 22.0
22:00 272.6 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 272.7 F 28.7
2:00 271.5 F 28.0
3:00 272.7 F 28.8
4:00 271.7 F 28.0

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.5 F 0.0
5:00 215.7 F 0.0
10:00 231.2 F 0.0
15:00 229.9 F 0.0
20:00 229.4 F 0.0
25:00 227.1 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 22:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

08/29/2022 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____