
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.6uS / 0.3ppm
CYCLE NUMBER 000043
11:28 21/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:53
136.4 C 224kPa 3:53
136.3 C 224kPa 4:23
136.2 C 224kPa 4:53
136.3 C 224kPa 5:23
136.2 C 224kPa 5:53
136.3 C 223kPa 6:23
136.2 C 224kPa 6:53
136.3 C 224kPa 7:23
136.3 C 224kPa 7:23
VENTING 7:23
AIR DRYING 8:14
CYCLE COMPLETE 23:14

Digital Signature #
B04524C36ECDA978
