
Midmark M11 - v1.0.5
Total Cycles: 1548

Name: RITTER2
Sterilizer ID: V2313228
09/29/2022 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.3 F 0.0
2:00 182.5 F 1.5
4:00 183.3 F 5.5
6:00 204.4 F 11.1
8:00 224.5 F 6.9
10:00 240.5 F 13.5
12:00 242.4 F 12.1
14:00 242.5 F 11.5
16:00 256.9 F 19.1
18:00 269.3 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.1 F 28.4
2:00 272.6 F 28.8
3:00 272.8 F 29.2
4:00 271.6 F 28.4

Min 271.5 F 28.2
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 227.1 F 0.0
5:00 231.2 F 0.0
10:00 224.8 F 0.0
15:00 228.1 F 0.0
20:00 226.1 F 0.0
25:00 218.1 F 0.0
30:00 214.4 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 18:55
STERILIZING: 4:00
VENTING: 2:41
DRYING: 30:00
TOTAL CYCLE: 00:57:38
CYCLE COMPLETE

09/29/2022 12:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____