
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001838
11:51 12/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:51
PRESSURIZING 1:57
STERILIZING 2:47
136.0 C 221kPa 2:47
136.2 C 225kPa 3:17
136.3 C 224kPa 3:47
136.4 C 225kPa 4:17
136.3 C 226kPa 4:47
136.3 C 224kPa 5:17
136.3 C 224kPa 5:47
136.1 C 221kPa 6:17
136.1 C 221kPa 6:17
VENTING 6:17
AIR DRYING 7:07
CYCLE COMPLETE 37:07

Digital Signature #
40A69C77B788BB80
