
Midmark M11 - v1.0.0
Total Cycles: 1153

Name: RITTER 1
Sterilizer ID: V2178198
09/09/2021 07:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 94.2 F 0.0
4:00 85.9 F 0.1
6:00 127.8 F 0.1
8:00 151.5 F 0.1
10:00 180.5 F 0.3
12:00 203.7 F 1.8
14:00 221.0 F 4.9
16:00 228.1 F 7.9
18:00 233.5 F 11.5
20:00 229.9 F 6.8
22:00 243.1 F 12.2
24:00 238.9 F 9.8
26:00 236.9 F 8.6
28:00 249.4 F 14.7
30:00 260.9 F 21.2
32:00 271.4 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.6 F 28.7
2:00 271.6 F 28.0
3:00 272.7 F 28.8
4:00 271.7 F 28.0

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.4 F 0.0
5:00 232.5 F 0.0
10:00 204.3 F 0.0
15:00 187.0 F 0.0
20:00 172.7 F 0.0
25:00 185.6 F 0.0
30:00 183.6 F 0.0

Drying Complete

FILLING: 1:11
HEATING: 32:40
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:10:55
CYCLE COMPLETE

09/09/2021 09:11 AM
ACCEPTED: X
REJECTED:
Operator: U02