
Midmark M11 - v1.0.5
Total Cycles: 2543

Name: RITTER2
Sterilizer ID: V2313228
06/21/2023 08:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.5 F 0.0
2:00 181.2 F 2.4
4:00 181.2 F 6.7
6:00 203.4 F 11.8
8:00 220.6 F 6.5
10:00 234.7 F 12.0
12:00 235.5 F 9.4
14:00 238.4 F 9.4
16:00 248.2 F 14.6
18:00 259.9 F 21.3
20:00 270.2 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.6 F 29.7
2:00 272.4 F 29.2
3:00 271.4 F 28.9
4:00 272.4 F 29.8

Min 271.4 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.4 F 0.0
5:00 235.9 F 0.0
10:00 226.4 F 0.0
15:00 224.1 F 0.0
20:00 223.7 F 0.0
25:00 229.3 F 0.0
30:00 230.2 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 20:53
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:37
CYCLE COMPLETE

06/21/2023 09:33 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____