

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5089

Name: RITTER 1  
Sterilizer ID: V2178198  
11/14/2024 09:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 71.3 F 0.0  
4:00 81.9 F 0.0  
6:00 107.3 F 0.0  
8:00 148.8 F 0.1  
10:00 170.8 F 0.5  
12:00 195.1 F 0.8  
14:00 211.9 F 2.4  
16:00 216.8 F 5.0  
18:00 221.5 F 7.7  
20:00 230.9 F 11.1  
22:00 232.2 F 7.4  
24:00 241.6 F 11.7  
26:00 237.7 F 9.4  
28:00 239.7 F 9.7  
30:00 248.2 F 13.9  
32:00 258.6 F 19.6  
34:00 268.0 F 25.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.4 F 28.6  
2:00 272.1 F 28.2  
3:00 271.4 F 27.9  
4:00 272.4 F 28.6  
  
Min 271.4 F 27.7  
Max 272.6 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.5 F 0.0  
5:00 222.2 F 0.0  
10:00 211.9 F 0.0  
15:00 213.5 F 0.0  
20:00 227.5 F 0.0  
25:00 230.6 F 0.0  
30:00 226.6 F 0.0  
Drying Complete

FILLING: 0:42  
HEATING: 35:31  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:14:07  
CYCLE COMPLETE

11/14/2024 11:13 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_