

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3821

Name: RITTER2  
Sterilizer ID: V2313228  
07/29/2024 11:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 134.1 F 0.0  
2:00 132.3 F 1.0  
4:00 134.3 F 3.0  
6:00 169.1 F 6.5  
8:00 193.2 F 10.5  
10:00 218.7 F 6.2  
12:00 223.1 F 7.3  
14:00 232.9 F 11.2  
16:00 229.7 F 6.7  
18:00 242.2 F 12.2  
20:00 238.9 F 10.3  
22:00 250.9 F 16.2  
24:00 261.7 F 22.3  
26:00 271.6 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.9  
1:00 272.2 F 29.0  
2:00 272.6 F 29.4  
3:00 271.5 F 29.2  
4:00 272.1 F 29.0

Min 271.2 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.2 F 0.0  
5:00 238.0 F 0.0  
10:00 229.8 F 0.0  
15:00 221.5 F 0.0  
20:00 218.0 F 0.0  
25:00 213.0 F 0.0  
30:00 192.9 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:33  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:05:04  
CYCLE COMPLETE

07/29/2024 12:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_