
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.1uS / 0.7ppm
CYCLE NUMBER 001487
11:45 10/07/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:39
PRESSURIZING 6:20
STERILIZING 7:52
136.0 C 220kPa 7:52
136.0 C 220kPa 8:23
136.4 C 224kPa 8:53
136.3 C 222kPa 9:23
136.4 C 223kPa 9:53
136.4 C 221kPa 10:23
136.3 C 222kPa 10:53
136.4 C 222kPa 11:23
VENTING 11:23
AIR DRYING 12:36
CYCLE COMPLETE 42:36

Digital Signature #
0335096E5C3D5670
