
Midmark M11 - v1.0.5
Total Cycles: 3129

Name: RITTER2
Sterilizer ID: V2313228
11/15/2023 03:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.0 F 0.0
2:00 140.7 F 1.5
4:00 162.8 F 4.6
6:00 184.0 F 8.8
8:00 207.2 F 13.2
10:00 216.6 F 7.2
12:00 229.0 F 12.0
14:00 232.0 F 8.7
16:00 239.9 F 11.1
18:00 245.0 F 13.3
20:00 256.9 F 19.7
22:00 267.9 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.1
1:00 271.8 F 29.7
2:00 272.4 F 29.6
3:00 271.3 F 29.0
4:00 272.1 F 30.0

Min 271.2 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 213.0 F 0.0
10:00 215.1 F 0.0
15:00 214.3 F 0.0
20:00 212.9 F 0.0
25:00 210.7 F 0.0
30:00 209.4 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 23:14
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:50
CYCLE COMPLETE

11/15/2023 04:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____