

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3211

Name: RITTER 1  
Sterilizer ID: V2178198  
05/10/2023 11:32 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 187.0 F 3.2  
4:00 193.9 F 7.4  
6:00 214.8 F 12.2  
8:00 224.2 F 6.7  
10:00 236.6 F 11.8  
12:00 235.2 F 8.8  
14:00 240.4 F 10.4  
16:00 246.4 F 13.3  
18:00 258.1 F 19.6  
20:00 268.5 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.8  
2:00 271.4 F 28.3  
3:00 272.4 F 28.9  
4:00 271.5 F 28.2  
  
Min 271.4 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.6 F 0.0  
5:00 218.3 F 0.0  
10:00 229.3 F 0.0  
15:00 230.3 F 0.0  
20:00 228.1 F 0.0  
25:00 226.2 F 0.0  
30:00 226.7 F 0.0  
Drying Complete

FILLING: 1:52  
HEATING: 21:13  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:00:22  
CYCLE COMPLETE

05/10/2023 12:32 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_