
Midmark M11 - v1.0.0
Total Cycles: 2928

Name: RITTER 1
Sterilizer ID: V2178198
02/22/2023 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 215.0 F 0.0
2:00 209.1 F 4.3
4:00 207.9 F 9.3
6:00 224.8 F 7.3
8:00 232.4 F 10.1
10:00 234.2 F 7.7
12:00 244.6 F 13.4
14:00 243.9 F 12.1
16:00 257.0 F 18.9
18:00 268.4 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.5 F 28.8
2:00 271.3 F 28.2
3:00 272.5 F 28.9
4:00 271.8 F 28.3

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.3 F 0.0
5:00 230.7 F 0.0
10:00 237.4 F 0.0
15:00 224.5 F 0.0
20:00 219.7 F 0.0
25:00 217.2 F 0.0
30:00 215.9 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 19:11
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

02/22/2023 11:56 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____