
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 001559
12:12 11/04/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:44
PRESSURIZING 5:47
STERILIZING 7:12
136.0 C 220kPa 7:12
136.3 C 222kPa 7:42
136.2 C 221kPa 8:12
136.5 C 224kPa 8:42
136.3 C 223kPa 9:12
136.4 C 220kPa 9:42
136.4 C 220kPa 10:12
136.4 C 223kPa 10:42
VENTING 10:42
AIR DRYING 11:45
CYCLE COMPLETE 41:45

Digital Signature #
C3282A2E809380EF
