

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3404

Name: RITTER2  
Sterilizer ID: V2313228  
02/22/2024 03:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.9 F 0.0  
2:00 156.4 F 2.4  
4:00 175.1 F 6.4  
6:00 198.8 F 11.4  
8:00 218.9 F 5.5  
10:00 233.7 F 10.5  
12:00 231.3 F 7.6  
14:00 243.9 F 13.6  
16:00 243.5 F 12.5  
18:00 256.2 F 19.3  
20:00 267.6 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.0  
1:00 272.6 F 30.0  
2:00 272.2 F 29.4  
3:00 271.1 F 28.9  
4:00 272.0 F 29.8  
  
Min 271.0 F 28.9  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.5 F 0.0  
5:00 235.5 F 0.0  
10:00 229.8 F 0.0  
15:00 226.3 F 0.0  
20:00 219.6 F 0.0  
25:00 231.8 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 21:26  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 00:59:53  
CYCLE COMPLETE

02/22/2024 04:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_