
Midmark M11 - v1.0.0
Total Cycles: 753

Name: RITTER 1
Sterilizer ID: V2178198
03/29/2021 08:23 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 147.3 F 0.8
4:00 168.9 F 3.2
6:00 192.7 F 8.2
8:00 210.5 F 11.1
10:00 221.2 F 7.9
12:00 237.5 F 13.2
14:00 237.4 F 10.4
16:00 233.5 F 7.9
18:00 249.8 F 14.8
20:00 261.6 F 21.5
22:00 272.4 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.0
1:00 272.7 F 28.7
2:00 272.6 F 28.7
3:00 272.4 F 28.8
4:00 272.5 F 28.9

Min 270.9 F 27.8
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 233.6 F 0.0
10:00 228.9 F 0.0
15:00 235.7 F 0.0
20:00 218.1 F 0.0
25:00 181.4 F 0.0
Dry Cycle Aborted

FILLING: 0:58
HEATING: 22:18
STERILIZING: 4:00
VENTING: 2:57
DRYING: 28:54
TOTAL CYCLE: 00:59:07

03/29/2021 09:22 AM
ACCEPTED: X
REJECTED:
Operator: U01