

-----  
Midmark M11 - v1.0.5  
Total Cycles: 243

Name: RITTER2  
Sterilizer ID: V2313228  
07/26/2021 08:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.2 F 0.0  
2:00 174.7 F 2.2  
4:00 176.6 F 6.2  
6:00 198.1 F 11.1  
8:00 215.3 F 5.3  
10:00 229.5 F 10.8  
12:00 231.1 F 8.1  
14:00 240.8 F 11.6  
16:00 245.0 F 13.3  
18:00 257.1 F 19.9  
20:00 268.4 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.1  
1:00 272.7 F 29.9  
2:00 271.6 F 29.1  
3:00 272.6 F 29.8  
4:00 271.6 F 29.0  
  
Min 271.5 F 29.0  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 233.7 F 0.0  
10:00 224.3 F 0.0  
15:00 220.3 F 0.0  
20:00 216.4 F 0.0  
25:00 214.6 F 0.0  
30:00 195.8 F 0.0  
Drying Complete

FILLING: 1:24  
HEATING: 21:06  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:59:40  
CYCLE COMPLETE

07/26/2021 09:04 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01