
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000789
16:38 10/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:55
PRESSURIZING 6:31
STERILIZING 8:15
136.4 C 223kPa 8:15
136.4 C 223kPa 8:45
136.2 C 222kPa 9:15
136.4 C 220kPa 9:45
136.3 C 222kPa 10:15
136.4 C 223kPa 10:45
136.1 C 221kPa 11:15
136.4 C 223kPa 11:45
136.4 C 223kPa 11:45
VENTING 11:45
CYCLE ABORTED 12:30
STOP BUTTON PRESSED
NOT STERILE
END TIME 12:30

Digital Signature #
B6EAD34CD4F92AE6
